Agricultural Laboratory 2004 Research Park Cir. Manhattan, KS 66502 Kansas Department of Agriculture

Phone: 785-564-7477 Fax: 785-564-6777 www.agriculture.ks.gov

Laura Kelly, Governor

K17242

4/12/2024

4/9/2024

4/11/2024

4/11/2025

Certificate Number :

Submission Date :

Calibration Date :

Due Date :

Certificate Issue Date :

Mike Beam, Secretary

Kansas Metrology Laboratory Calibration Certificate

Submitted by:	Coastal Rail Scales LLC
	7834 Lake Street
	Lake Charles, LA 60705

Condition : used

	Calibrated	Adjusted	Rejected	Temperature		Rela mperature Hum	tive idity
I	1	0	0	20.1 °C	4	20.1 °C 55.	5 %
				20.2 °C		20.2 °C 55.	)%

Conformity Statement: Standards are evaluated for conformance to specifications, except for density, magnetism, and surface finish. Assumed density, which is obtained from documentation about the standards, is used where applicable. Standards are calibrated as found, after acclimation with laboratory conditions, and recalibrated if adjustment is required. Standards conform to specifications and maximum permissible errors (MPE), unless listed as "Rejected" or otherwise noted. Specifications, MPE's, or their sources listed as "N/A" are not applicable to the calibration. Measurements and uncertainties comply with published decision rules that require the expanded uncertainty to be less than 1/3 the MPE and require the measurement plus or minus the expanded uncertainty to be within the MPE. True mass is the weight of a standard at a reference air density of 0.0012 g/cm<sup>3</sup>, compared to reference standards at a reference density of 8.0 g/cm<sup>3</sup>, at a reference temperature of 20 °C.

Calibration Procedures: • For weight carts: NISTIR 6969 (May 2019): SOP 33

Specifications and MPE Sources:

• For weight carts: NISTHB 105-8 (May 2019)

Conversion Factors: 1 kg = 1 000 g = 1 000 000 mg, 1 lb = 453.592 37 g, 1 oz = 28.349 523 125 g, 1 lb = 16 oz

Uncertainty Statement: The combined standard uncertainty includes the standard uncertainty reported for the standards, the standard uncertainty for the measurement process, the standard uncertainty for any uncorrected errors associated with buoyancy corrections (applies to mass values only), the standard uncertainty for any uncorrected errors associated with temperature correction (applies to length and volume values only), the standard uncertainty for balance drift, the standard uncertainty of balance sensitivity, and a component of uncertainty to account for any observed deviations from the National Institute of Standards and Technology (NIST) values that are less than surveillance limits. The combined standard uncertainty is multiplied by the coverage factor (k-value), which is based on the degrees of freedom, to give the expanded uncertainty, which defines an interval having a level of confidence of 95.45 %. The expanded uncertainty contained in this certificate is consistent with the OIML G 1-100 Edition 2008 (E), Evaluation of measurement data – Guide to the expression of uncertainty in measurement. The expanded uncertainty is not to be confused with the MPE during application.

Traceability Statement: The standards of the Kansas Metrology Laboratory are traceable to the International System of Units (SI) through NIST and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The certificate number listed above is the unique identifier to be used when referencing measurement traceability for the items contained in this certificate only.

This document does not represent or imply endorsement by the Kansas Metrology Laboratory, NIST, or any state or federal government agency. The data in this certificate only applies to those items specifically listed in this certificate. This document may not be reproduced, except in full, without the written permission of the Kansas Metrology Laboratory.

Kevin Uphoff, State Metrologist

Certificate Number: K17242

Nominal Mass	Serial Number	Conventional Mass as Found	Standards Class	MPE ±	Expanded Uncertainty (U) ±	Coverage Factor (k)	Conventional Mass as Left	Adjusted/ Within MPE/ Rejected
8000 lb	WC2015001	3629070 g	NIST HB 105-8	1270 g	300 g	2.04	3629070 g	Within MPE